

Performance Exam Checklist

Nondestructive Measurement of Thickness of Nonmagnetic Coatings on a Ferrous Base FOP For ASTM D 7091

Technician Name _____ Exam Date _____

Procedure Element	Yes	No
1. The tester has a copy of the current procedure on hand?	<input type="checkbox"/>	<input type="checkbox"/>
2. All equipment is functioning according to the test procedure, and if required, has the current calibration/verification tags present?	<input type="checkbox"/>	<input type="checkbox"/>
3. Instrument calibrated in accordance with the manufacturer's instructions before use employing a suitable thickness standard?	<input type="checkbox"/>	<input type="checkbox"/>
4. Several readings taken and recorded taking into account edge and curvature effects?	<input type="checkbox"/>	<input type="checkbox"/>
5. The average thickness converted to oz. ft ² (g/m ²) using appropriate conversion factor?	<input type="checkbox"/>	<input type="checkbox"/>

First attempt: Pass Fail

Second attempt: Pass Fail

Signature of Examiner _____

Comments:

